




PFE ≥ 90%



Child Mask
PFE ≥ 95%

XING GUANGTM

Shandong XingGuang Protective Products Co.,Ltd.

 Zhilan village 3, Gaomi Economic Development Zone, Gaomi City,
Shandong Province, P.R.China(261500)

 www.xingguang-protective.com






Specification:

Features:

- Face and head comfort, Skin comfort, Lightweight
- PFE $\geq 90\%$
- Non-Sterilization



Material	First layer-- PP non-woven 30g,3S
	Second layer--25g Meltblown non-woven 99% filter efficiency
	Third layer--PP non-woven 30g,3S
Style	Elastic Ear-Loop , Ties
Aluminum nose clip	Metal(Inner)
Certificates	CE
Standard	GB/T32610-2016, EN 14683 Type I Pending
PFE	$\geq 90\%$
Size	17.5x9.5CM
HS Code	6307900010



DISPOSABLE CHILD MASK

Features:

- Face and head comfort, Skin comfort, Lightweight
- PFE $\geq 95\%$
- Non-Sterilization



Specification:

Material	First layer-- PP non-woven 30g,3S
	Second layer--25g Meltblown non-woven 99% filter efficiency
	Third layer--PP non-woven 30g,3S
Style	Elastic Ear-Loop, Ties
Aluminum nose clip	Metal(Inner)
Certificates	CE
Standard	GB/T32610-2016, EN 14683 Type I Pending
PFE	$\geq 95\%$
Size	14.5 X 9.5cm
HS Code	6307900010

Packing Manner:

10pcs /polybag,50pcs/box,2,000pcs/carton
 size:520x420x420mm
 GW:7.5KGS

10PCS/polybag



50PCS/box



2000PCS/carton







Loading Container by Sea:

- 20FT Container:250 cartons/550,000pcs
- 40HQ Container:550 cartons/1,200,000pcs



Test Report by CNAS approved lab




佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

Testing Report	Security website: www.fcl-sz.org.cn Security code: 7215361000	
Report No: ZFLJ2619901A	Page 1 of 3	

Applicant Information

Applicant Name : Shandong Xingguang Protective Products Co.,Ltd.
 Applicant Address : Zhilan village 3,Gaomi Economic Development Zone,Gaomi City 261500
 Manufacturer : Shandong Xingguang Protective Products Co.,Ltd.

Sample Information

Sample Description : Non-medical disposable protective mask
 Style No. : XG-1
 Sample Quantity : 50 pieces
 Colour : Blue
 Size specification : 9.5X17cm


- Sample Receiving Date : 2020-05-08
 - Report Date : 2020-05-18
 - The original sample is stuck on the last paper.


Test Performed

Judgement according to:
 GB/T 32610-2016 Technical Specification of daily protective mask
 - Selected test(s) as requested by applicants. For details, refer to attached page(s).




Pronounce


The results shown in this report refer only to sample(s) tested unless otherwise stated. All the items are performed in the standard conditions, except the noted cases. Except for the requirement of the client, the test results and the conformity judgement of this report do not take the uncertainty of the test results into account.

Signed for and on behalf of
CNTAC Testing Service Co.,Ltd.(Foshan)
 Approved by




广东省佛山市南海区西樵镇南方技术创新中心一、三、四楼1,3&4F., Nanfang Technology Innovation Center, Xiqiao,Nanhai District, Foshan City, Guangdong, China
 Tel: (86 757)86850633/86806656 Fax: 86850633 Http://www.fcl.org.cn
 若对检测报告有异议, 应于收到报告之日起 15 日内向检测单位提出, 逾期不予受理。





佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

Testing Report	Security website: www.fcl-sz.org.cn Security code: 7215361000	
Report No: ZFLJ2619901A	Page 2 of 3	




--- Test Result ---


Test Result	Requirements	Judgement
1. Filtration Efficiency(Salt medium) GB/T 32610-2016 Appendix A		
Air flow: 85L/min, Aerosol: NaCl Unit: <math>\%>		
Minimum value	95.2	Level III ≥ 90 Pass
2. Exhalation Resistance GB/T 32610-2016 Section 6.8		
Air flow: 85L/min Unit: <math>\text{Pa}>		
Maximum value	42.4	≤ 145 Pass
3. Inhalation Resistance GB/T 32610-2016 Section 6.7		
Air flow: 85L/min Unit: <math>\text{Pa}>		
Maximum value	44.0	≤ 175 Pass
4. Appearance Requirement GB/T 32610-2016 section 6.1		
Result	Accordance	Mask surface should not be damaged, oil stains, deformation and other obvious defects. Pass

Remark: 1.The inhalation resistance and exhalation resistance items are unauthorized for CNAS.
 2. This report is a supplement to the report number ZFLJ2619901.




广东省佛山市南海区西樵镇南方技术创新中心一、三、四楼1,3&4F., Nanfang Technology Innovation Center, Xiqiao,Nanhai District, Foshan City, Guangdong, China
 Tel: (86 757)86850633/86806656 Fax: 86850633 Http://www.fcl.org.cn
 若对检测报告有异议, 应于收到报告之日起 15 日内向检测单位提出, 逾期不予受理。





佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

Testing Report	Security website: www.fcl-sz.org.cn Security code: 7215361000	
Report No: ZFLJ2619901A	Page 3 of 3	

Original Sample



***** End of Report *****



广东省佛山市南海区西樵镇南方技术创新中心一、三、四楼1,3&4F., Nanfang Technology Innovation Center, Xiqiao,Nanhai District, Foshan City, Guangdong, China
 Tel: (86 757)86850633/86806656 Fax: 86850633 Http://www.fcl.org.cn
 若对检测报告有异议, 应于收到报告之日起 15 日内向检测单位提出, 逾期不予受理。

NMPA Regisrtation

2020/5/20 国家药品监督管理局-数据查询 中文 | Engl

国家药品监督管理局
National Medical Products Administration

机构概况 政务公开 药品 医疗器械 化妆品

网站首页 >> 数据查询 >> 国产医疗器械产品 (备案)

产品分类: 医疗器械

高级查询

医疗器械 备案号: 鲁潍械备20200095号

备案人名称: 山东星光防护用品有限公司

备案人注册地址: 山东省潍坊市高密市高密经济开发区芝兰三村315号营业房

生产地址: 山东省潍坊市高密市经济开发区芝兰一村横二路兆农田东西100米, 南北100米

产品名称: 医用隔离面罩

型号规格: 普通型 耳挂式

产品描述: 通常由高分子材料制成的防护罩、泡沫条和固定装置组成。非无菌提供, 一次性使用。

预期用途: 用于医疗机构中检查治疗时起防护作用, 阻隔体液、血液飞溅或泼溅。

备注: 备案单位 山东省食品药品监督管理局 备案日期 2020-04-20 变更情况

产品存储及有效期

注: 网站发布的医疗器械注册和产品备案信息供公众查询, 如网站公布的信息与图纸度文件不一致, 请联系相应部门补修。进口和国产三类医疗器械数据问题请联系国家药监局医疗器械数据注册处: qijieju@nmpa.org.cn【邮件主题请注明“医疗器械数据问题”】。邮件正文中需详细填写以下全部信息: 医疗器械注册证号/备案号; 2.类型(注册、变更、延续等); 3.问题描述(500字以内); 4.企业名称(全称); 5.统一社会信用代码; 6.联系人姓名; 7.联系电话(手机和座机); 8.联系邮箱; 或致电国家药监局医疗器械数据注册处电话010-88331951【此电话为我局医疗器械数据注册联系电话, 并非我局产品/企业业务咨询电话】。国产一类和二类医疗器械数据问题请联系企业所在地药监, 由省局通过数据共享平台进行协调和维护。

中国政务网 国家药监局 国家市场监督管理总局 直属单位 地方药监局

本局由国家药品监督管理局主办, 版权所有。国家药品监督管理局 网站声明
Copyright © NMPA All Rights Reserved
网站标识码bm15000001 备案序号: 鲁ICP备13027807号 京公网安备11010202068311号
地址: 北京市西城区药监局北大街1号 | 邮编: 100037 | 固定电话: 68311166

第一类医疗器械备案信息表

备案号: 鲁潍械备 20200095 号

备案人名称	山东星光防护用品有限公司
备案人组织机构代码	(境内医疗器械通用)
备案人注册地址	山东省潍坊市高密市高密经济开发区芝兰三村315号营业房
生产地址	山东省潍坊高密市经济开发区芝兰一村横二路兆农田东西100米, 南北100米
代理人	(进口医疗器械通用)
代理人注册地址	(进口医疗器械通用)
产品名称	医用隔离面罩
型号/规格	普通型 耳挂式
产品描述	通常由高分子材料制成的防护罩、泡沫条和固定装置组成。非无菌提供, 一次性使用。
预期用途	用于医疗机构中检查治疗时起防护作用, 阻隔体液、血液飞溅或泼溅。
备注	
备案单位	
和日期	备案日期: 2020年04月20日
变更情况	

第一类医疗器械备案凭证

山东星光防护用品有限公司:

根据相关法规要求, 对你单位第一类医疗器械: 医用隔离面罩予以备案, 备案号: 鲁潍械备 20200095 号

日期: 2020年04月20日

潍坊市市场监督管理局

Our factory

Production capacity:

KN95 200,000pcs/Day

3ply Disposable mask 500,000pcs/Day



•Description:
Factory



•Description:
Production Lines



•Description:
Export License



•Description:
Medical Device Registration



•Description:
Factory



•Description:
Office



•Description:
Business License

•Description:
Medical Device Registration



•Description:
Medical Device Registration